


# Product Specifications

## FlashLink Certified Vaccine USB Data Logger

<b>Measurement Range</b>	-40°F to 104°F (-40°C to 40°C)
<b>Display Resolution</b>	0.1°F/°C
<b>Accuracy within Normal Operating Range</b>	±1.0°F from 14°F to 77°F (±0.5°C from -10°C to 25°C)
<b>Data Capacity</b>	32,496 data points
<b>Data Download Time</b>	2 seconds typical for 3823 points
<b>Programmable Alarm Levels</b>	High & Low levels over entire range, displayed on LCD
<b>Battery</b>	Coin Cell Lithium battery, % remaining power indicator on LCD
<b>Dimensions/Body</b>	1.575" x 3.189" x .575" (40mm x 81mm x 15mm)
<b>Logger Casing</b>	ABS plastic
<b>External Sensor</b>	5mL PETG* sterile bottle – 21CFR compliance information available, 9.84 ft./3m cable
<b>Glycol Buffer</b>	Propylene glycol (GRAS** classified)
<b>Weight</b>	2.56 oz. (70 grams)
<b>Certification</b>	CE Mark, NIST traceable, RoHS Compliant
<b>Display</b>	LCD, Red LED Alarm
<b>Waterproof Rating</b>	IP64 (with silicone cap)
<b>Storage conditions</b>	-4°F to 122°F (-20°C to 50°C) 0% to 95% RH w/o condensation
<b>Shelf Life</b>	2 Years
<b>System Requirements</b>	Windows PC with current operating system, Adobe Reader
<b>Programming</b>	FlashPDF Program Manager software
<b>Download Requirements</b>	Direct USB connection to PC
<p>*Polyethylene terephthalate copolymer                  **Generally Recognized As Safe</p>	



Model	
<b>20932</b>	Includes Manufacturers Certificate of Compliance traceable to NIST
<b>20938</b>	Includes ISO 17025 Calibration Certificate from A2LA accredited lab, valid for 2 years

- WHO prequalified under PQS Code E006/026
- Compatible with FDA and CDC guidelines
- 21 CFR Part 11 compliant
- Single external temperature sensor
- Glycol buffered probe emulates products and prevents false alarms
- Ideal for monitoring storage refrigerators and freezers
- User programmable: logging intervals, hi/low alarms
- Large, easy to read LCD
- Red flashing LED alarm for temperature excursions
- Low battery indicator on LCD

DeltaTrak manufactures products under an ISO 9001:2008 registered quality management system

SS0141\_16C2

